

This is the author's accepted manuscript of an article published in *Journal of Feline Medicine and Surgery*.

The final publication is available at SAGE Journals via <https://doi.org/10.1177%2F1098612X18813426>.

The full details of the published version of the article are as follows:

TITLE: Inter-rater and inter-device reliability of mechanical thresholds measurement with the Electronic von Frey Anaesthesiometer and the SMALGO in healthy cats

AUTHORS: Adami, C; Lardone, E; Monticelli, P

JOURNAL TITLE: Journal of Feline Medicine and Surgery

PUBLICATION DATE: 22 November 2018 (online)

PUBLISHER: SAGE Publications

DOI: <https://doi.org/10.1177%2F1098612X18813426>

1 **Original article**

2 **Inter-rater and inter-device reliability of mechanical thresholds measurement with**  
3 **the Electronic von Frey Anaesthesiometer and the SMALGO in healthy cats**

4

5 **Chiara Adami<sup>1</sup>, Elena Lardone<sup>2</sup> and Paolo Monticelli<sup>1</sup>**

6 <sup>1</sup>Department of Clinical Sciences and Services, Royal Veterinary College, University of  
7 London, UK

8 <sup>2</sup>Department of Veterinary Science, University of Turin, Italy

9

10 **Corresponding author:**

11 Chiara Adami DMV, MRCVS, DACVAA, DECVAA, RCVS and EBVS® □ European  
12 Specialist in Anaesthesia and Analgesia, PhD, Department of Clinical Sciences and  
13 Services, Royal Veterinary College, Hawkshead Lane, AL97TA, Hatfield, UK

14 Email: [cadami@rvc.ac.uk](mailto:cadami@rvc.ac.uk)

15

16

17

18

19

20

21

22 **Abstract**

23 **Objectives** To compare the Electronic von Frey Anaesthesiometer (EVF) and the Small  
24 Animal ALGOmeter (SMALGO), used to measure sensory thresholds in 13 healthy cats  
25 at both the stifle and the lumbosacral joint, in terms of inter-rater and inter-device  
26 reliability.

27 **Methods** Two independent observers carried out the sets of measurements in a  
28 randomised order, with a 45-minute interval between them, in each cat. The inter-rater  
29 and inter-device reliability were evaluated by calculating the inter-rater correlation  
30 coefficients (ICC) for each pair of measurements. The Bland-Altman method was used  
31 as an additional tool to assess the level of agreement between the two algometers.

32 **Results** The sensory thresholds measured with the EVF were  $311 \pm 116$  g and  $378 \pm 178$   
33 g for the stifle and for the lumbosacral junction, respectively, whereas those measured  
34 with the SMALGO were  $391 \pm 172$  g and  $476 \pm 172$  g. The inter-rater reliability was fair  
35 (ICC > 0.4) for each pair of measurements except those taken at the level of the stifle with  
36 the SMALGO, for which the level of agreement between observer A and B was poor (ICC  
37 = 0.01). The inter-device reliability was good (ICC = 0.73; P= 0.001). The repetition of  
38 the measurements affected reliability, as the thresholds obtained after the 45 minute break  
39 were consistently lower than those measured during the first part of the trial (P = 0.02).

40 **Conclusions and relevance** The EVF and the SMALGO may be used interchangeably in  
41 cats, especially when the area to be tested is the lumbosacral joint. However, when the  
42 thresholds are measured at the stifle, the inter-observer reliability is better with the EVF

43 than with the SMALGO. The reliability decreases when the measurements are repeated  
44 within a short time interval, suggesting a limited clinical applicability of quantitative  
45 sensory testing with both algometers in cats.

46

47

48

49

50

51

52

53

54

55

56

57

58

59

60

61

62

63

64 **Introduction**

65 Recognizing and treating pain in feline patients has always been extraordinarily  
66 challenging. Traditionally, behavioural indicators are used to evaluate pain in cats,<sup>1,2</sup> and  
67 various species-specific pain scales have been developed on the basis of such indicators  
68 with the purpose of ameliorating peri-operative pain management.<sup>3,4</sup> Recently, the use of  
69 facial expressions as an additional tool to assess acute pain has become popular also in  
70 feline patients.<sup>5</sup>

71 Whilst for the evaluation of peri-operative acute pain the veterinarians can rely on  
72 a number of available and validated tools, scoring chronic pain remains a challenge even  
73 for the most experienced observers. Despite the lack of a unanimously accepted  
74 characterization of chronic pain in cats,<sup>6</sup> as a matter of fact cats do suffer from clinical  
75 conditions, such as osteoarthritis (OA),<sup>7</sup> which in humans and dogs are known to cause  
76 maladaptive pain.<sup>8-10</sup> In an attempt to evaluate OA-related feline pain, Benito and  
77 colleagues<sup>11</sup> developed and validated a Feline Musculoskeletal Pain Index (FMPI), based  
78 on subjective assessments performed by the owner in the animals' natural environment.  
79 With the same purpose, another study proposed the combined use of more objective  
80 parameters, namely gait analysis variables and mechanical sensory thresholds measured  
81 with an algometer.<sup>12</sup> Similarly, the Montreal Instrument for Cat Arthritis Testing,  
82 developed by Klinck and colleagues,<sup>13</sup> relies on a combination of behavioural indicators,  
83 mechanical thresholds and gait analysis.

84           The use of mechanical sensory thresholds as a tool to quantify chronic pain in cats  
85 is not novel, with most of the previous investigations that focused on this aspect relying  
86 on the use of the Electronic von Frey Anaesthesiometer (EVF).<sup>13-15</sup> This algometer is  
87 composed of a control unit and a sensory probe, used to apply over the body surface a  
88 force that is measured, displayed and stored. The force at which a predefined behavioural  
89 response is evoked is defined as threshold. Whilst the EVF has been designed for use in  
90 human patients, the Small Animal ALGOmeter (SMALGO), which shares with the  
91 former the working principle, has been specifically developed for laboratory rodents, and  
92 may represent a valid alternative to the EVF. The SMALGO was found useful and reliable  
93 to quantify pain in rats and mice in various experimental models, including inflammatory  
94 pain, mechanical allodynia and hyperalgesia.<sup>16-18</sup>

95           The primary aim of this study was to compare the EVF and the SMALGO, used to  
96 measure mechanical sensory thresholds in a population of healthy cats, at two anatomical  
97 sites commonly affected by feline OA, in terms of inter-rater and inter-device reliability.  
98 Secondary aims were to determine the effect of the repetition of a whole set of  
99 measurements, after a 45 minute-interval, on the reliability of both algometers, and to  
100 determine baseline mechanical sensory thresholds in healthy cats.

101           The authors hypothesized that the EVF and the SMALGO would be comparable for  
102 the use intended in this study, and that both inter-rater and inter-device reliability would  
103 be fair.

104

105 **Materials and methods**

106

107 **Ethical approval**

108 The study was conducted under ethical approval of both the University of Turin (Protocol  
109 number: 1245/120618) and the Clinical Research Ethical Review Board of the Royal  
110 Veterinary College of the University of London (License number: URN 2018 1773-3). A  
111 signed informed owner consent was obtained for each cat.

112

113 **Animals**

114 Thirteen cats, owned by either veterinarians or students in their 5<sup>th</sup> year of veterinary  
115 medicine, were enrolled in this study. The sample size was determined with the method  
116 described by Walter and colleagues (1998) for reliability studies, with the variables set as  
117 follows: number of observers = 2; desired value for inter-class correlation coefficient  
118 (ICC) = 0.8; minimally acceptable value for ICC = 0.05;  $\alpha = 0.05$  and  $\beta = 0.2$ . This  
119 resulted in a minimal number of observations (cats) equal to 10. The Exclusion criteria  
120 were history of orthopaedic and neurological conditions that may have altered the sensory  
121 thresholds, and medical therapy with any drug with known analgesic effect. The cats were  
122 admitted to the Veterinary teaching Hospital of the University of Turin on the morning  
123 of the data collection, and left undisturbed for acclimatization in the examination room  
124 where the measurements were carried out for at least 15 minutes. Demographic data  
125 collected and used for statistics were sex, breed, age (months), body Condition Score

126 (BCS: 0-9),<sup>20</sup> body weight (kg) and height (cm), the latter measured from the dorsal end  
127 of the scapular spine, identified by palpation, to the surface of the examination table, with  
128 the cat in standing position. Food and water remained available until the trial was  
129 commenced.

130

### 131 **Preparation of the instruments**

132 Both devices are calibrated at the factory and do not require recalibration prior to use.  
133 However, before each set of measurements, the EVF was checked for accuracy as follows.  
134 After the 1000 g probe was equipped with a new rigid tip, a standard 5.3 g weight  
135 provided by the manufacturer was applied onto the tip, with the unit in horizontal position.  
136 The measurements were allowed to begin only in case the reading displayed and stored  
137 by the unit was equal to  $5.3 \pm 0.1$  g. Regarding the SMALGO, the probe was equipped  
138 with the 3 mm sensor tip and the unit selected (g); following, the control unit was zeroed  
139 by resetting the tar to zero with a foot switch, and the key “max” pressed to allow the  
140 device to store the maximum force value recorded during the measurement.

141

### 142 **Sensory thresholds measurements**

143 Two anatomical sites were investigated: the lumbosacral intervertebral joint and the  
144 medial site of the stifle. The former was identified by using as anatomical landmarks the  
145 ileum wings, the last lumbar vertebra and the sacrum. For the latter, the target was the  
146 medial aspect of the knee, between the patella (dorsal) and the tibial tuberosity (ventral).



147 For both sites, the sensory tips of both instruments were applied perpendicularly to the  
148 skin, and a steadily increasing force applied until a positive behavioural reaction could be  
149 evoked. Attempt to escape, tail wiggling, hissing, attempts to bite or aggressions, ears  
150 back and flat against the head, head turning towards the site of stimulation, back muscle  
151 contraction (for the lumbosacral) and limb withdrawal (for the stifle) were considered  
152 positive behavioural reactions. When at least one of these reactions was observed, the  
153 mechanical stimulation was interrupted and the sensory tip withheld; the maximal force  
154 value displayed by the control unit was manually recorded. Each single measurement was  
155 repeated once to confirm the threshold, with a time interval of at least 30 seconds in order  
156 to avoid temporal summation;<sup>21</sup> the average calculated from these values was used for  
157 statistical analysis. Two observers (EL, observer A and CA, observer B) carried out the  
158 measurements independently, with the cats minimally restrained by the owner. A 45  
159 minute-time interval was allowed between the subsequent sets of measurements carried  
160 out by the two observers. For each cat, the order of the observers and, for each observer,  
161 of the device to be used first and of the anatomical site to be assessed first, was determined  
162 by simple randomization based on flipping of a coin.

163

#### 164 **Statistics**

165 Data distribution was assessed with both the Kolmogorov-Smirnov test and the Shapiro-  
166 Wilk test. The Spearman correlation coefficient (SCC) was calculated to detect  
167 correlations between the sensory thresholds and demographic variables of the cats (age,

168 BCS, body weight and height). The inter-observer reliability was evaluated between  
169 observers A and B and, for each cat, between the observer who started the trial (1<sup>st</sup>  
170 observer) and the other, who carried out the measurements after the 45 minute break (2<sup>nd</sup>  
171 observer). The levels of agreement were quantified by calculating the inter-rater  
172 correlation coefficient (ICC), with 95% confidence intervals (CI; upper and lower  
173 bounds). The inter-device reliability was evaluated with both the ICC (with CI) and the  
174 Bland-Altman analysis.<sup>22</sup> A paired-T test was run to compare sets of measurements  
175 showings means and standard deviations that appeared to be different at first sight  
176 (between observer A and B, and between the 1<sup>st</sup> and the 2<sup>nd</sup> observer). *P* values < 0.05  
177 were considered statistically significant. The level of agreement (both inter-observer and  
178 inter-device) was scored as follows: ICC < 0.40= poor; ICC between 0.40 and 0.59= fair;  
179 ICC between 0.60 and 0,74= good; and ICC between 0.75 and 1= excellent.<sup>23</sup>  
180 Commercially available softwares were used (IBM SPSS Statistics 24, IBM Corporation,  
181 NY, USA; and SigmaPlot 14 and SigmaStat 4, SYSTAT Software Inc, CA, USA).

182

### 183 **Results**

184 Normally distributed data are here presented as means and standard deviations, while data  
185 with non-normal distribution as medians and maximum-minimum ranges.

186 Twelve cats completed the study. One cat appeared to be stressed after the first set  
187 of measurements with the SMALGO, therefore it was decided to let him rest for about a

188 hour and then allow the second observer to proceed with the measurements only with the  
189 SMALGO, in order to use these two sets of data for comparison.

190 Five cats were spayed females while the remaining 8 cats were neutered males. The  
191 represented breeds were domestic short hair (n=12) and domestic long hair (n=1). The  
192 cats were aged 60 [12-180] months, weighed  $5.4 \pm 1.2$  kg, had a BCS of 5 [4-9] and their  
193 height was  $28 \pm 3.6$  cm. There were significant positive correlations between both the  
194 body weight and the BCS, and the sensory thresholds (SCC: 0.21 and 0.27; and  $P = 0.04$   
195 and 0.007, respectively), and significant negative correlation between the height of the  
196 cats and the sensory thresholds (SCC: -0.31;  $P = 0.001$ ). No correlation was found  
197 between the age of the cats and their sensory thresholds.

198 Observer A carried out the first set of measurements in 8 cats, while observer B  
199 started the trial in the remaining 5. There were no statistically significant differences  
200 between the sensory thresholds recorded by observers A and B, with both devices and at  
201 both anatomical sites. Overall, the thresholds recorded during the first set of  
202 measurements by one of the two observers (1<sup>st</sup> observer) with both devices and at both  
203 sites were significantly higher than those carried out by the other observer after the 45  
204 minute break (2<sup>nd</sup> observer) (Table 1;  $P = 0.02$ ). The level of agreement between these  
205 sets of measurements was poor (Table 2). The overall inter-rater agreement between 1<sup>st</sup>  
206 and 2<sup>nd</sup> observer was poor; however, when investigated in details, such agreement was  
207 fair when the measurements were carried out with the EVF at both the anatomical sites,  
208 and with the SMALGO at the lumbosacral joint, but poor for the measurements obtained

209 with the SMALGO at the stifle (Table 2). The inter-device reliability was good (Table 2;  
210 P value: 0.001), although the level of agreement between the EVF and the SMALGO was  
211 better at the lumbosacral junction compared to the stifle, as demonstrated by the higher  
212 ICC obtained at the former site (Table 2). Data for each variable are presented in Table  
213 1; the ICC for each set of comparison, together with the corresponding 95% CI, are shown  
214 in Table 2.

215

216

217

218

219

220

221

222

223

224

225

226

227

228

229

230 **Figure legend**

231 **Figure 1** The Bland-Altman plot shows the difference between the thresholds measured  
232 with the EVF and those with the SMALGO (g) in 13 healthy cats, plotted against the  
233 average of all the measured thresholds.

234

235

236

237

238

239

240

241

242

243

244

245

246

247

248

249

250

251 **Discussion**

252 This study demonstrates that the measurement of sensory thresholds in healthy cats with  
253 both the SMALGO and the EVF does not result in consistent readings when the  
254 measurements are repeated after a relatively short time interval. In each cat, the repetition  
255 of the trial 45 minutes after the first set of measurements resulted in decreased sensory  
256 thresholds, which seems to indicate that the cats easily became sensitized or less  
257 cooperative after manipulation. Since a useful method to quantify pain should be  
258 repeatable in order to evaluate the efficacy of the analgesic therapy and titrate it to effect,  
259 this drawback limits the clinical applicability of quantitative sensory thresholds in feline  
260 patients. It also suggests that, if repeated tests are to be performed, a time interval longer  
261 than 45 minutes between subsequent measurements may help to improve reliability.

262 The good inter-device reliability indicates that the thresholds measured with the two  
263 algometers are similar, and suggests that both the EVF and the SMALGO might be used  
264 interchangeably in cats. However, comparable results are more likely to be obtained when  
265 the two algometers are used to measure sensory thresholds at the lumbosacral junction  
266 than at the level of the stifle. Moreover, both observers obtained higher thresholds with  
267 the SMALGO compared to the EVF. A possible explanation for this finding could be that  
268 the 3 mm sensory tip, chosen by the authors for the SMALGO, is too small for cats and  
269 needs a greater application force than the EVF probe to evoke comparable behavioural  
270 reactions. The 3 mm tip was chosen over the 5 and 8 mm ones as our clinical experience

271 suggested that the former, owing to the pointed tip that applies the force on a small surface  
272 area, would evoke more consistent reactions than the flat 5 and 8 mm tips in cats.

273 Although the overall inter-rater agreement was poor, when this variable was  
274 analysed in details it showed that the agreement between observer A and observer B was  
275 fair for all pairs of measurements except the ones taken at the stifle with the SMALGO.  
276 The very poor agreement of this single comparison significantly affected the overall inter-  
277 rater agreement calculated between observer A and observer B, and could have been  
278 caused by a number of factors, including inappropriate selection of the SMALGO sensory  
279 tip, of the anatomical site, or both.

280 Investigating the feasibility of sensory thresholds as possible clinical tool to  
281 quantify, in the next future, pain in cats with degenerative joint disease was one of the  
282 focuses of this study. As a result, the stifle and the lumbosacral joint were chosen by the  
283 authors as anatomical sites of interest owing to their common involvement in feline  
284 osteoarthritis.<sup>15,24,25</sup> However, both investigators found the feline stifle a challenging  
285 anatomical site in terms of approachability when the cats were standing, and consistency  
286 and repeatability of the positioning of the sensory tip and subsequent application of the  
287 force. Regarding the future use of the EVF and of the SMALGO in the clinical setting, it  
288 is worth to consider that one of the intrinsic limitations of the current study is that its  
289 findings do not allow any conclusive statement about the validity of both devices for  
290 measuring pain in cats with actual OA.

291            Interestingly, physical variables of the cats, such as the height, the body weight and  
292 the BCS, had an effect on the sensory thresholds, which were higher in fat and heavier  
293 cats, and lower in taller, larger cats. Whilst the former finding could be due to the  
294 dampening effect of the adipose tissue covering both the lumbosacral joint and the stifle,  
295 which could have increase the tolerance of the cats to the mechanical stimulation in the  
296 area, providing a reasonable explanation for the inverse relationship between height and  
297 sensory threshold is more challenging. It might be hypothesized that large sized cats are  
298 more prone to develop osteoarthritis owing to the increased load on the joints, and that  
299 some of the taller cats of this study were affected. One study found that large breed cats,  
300 such as Maine coon, are prone to develop hip dysplasia.<sup>26</sup> However, whilst obesity and  
301 elderly are recognized risk factor for feline OA,<sup>27</sup> there is no published evidence that the  
302 size of the cats may act as well as predisposing condition. On the other hand, in this  
303 current study fatter cats had higher sensory thresholds, which indicates a higher tolerance  
304 to mechanical stimulation, and no correlation was found between sensory thresholds and  
305 elderly. The cats of the current study were owned by either a veterinarian or a veterinary  
306 medicine student, and regularly underwent clinical exam on occasion of standard  
307 vaccinations and deworming. Moreover, all owners were caring to their cats and it is  
308 reasonable to assume that they would notice changes in behaviour or signs of severe pain.  
309 Nevertheless, owing to the lack of a thorough orthopaedic and radiographic examination,  
310 the presence of osteoarthritis cannot be ruled out.

311



312 **Conclusions**

313 The good inter-device reliability suggests that the EVF and the SMALGO may be used  
314 interchangeably in cats; nevertheless, the poor inter-rater reliability observed when the  
315 SMALGO was used at the stifle indicates that, for this anatomical site, the EVF may  
316 represent a better option. Repetition of the measurements within a short time interval does  
317 affect reliability, a drawback that may limit the applicability of quantitative sensory  
318 testing with both algometers in clinical feline patients.

319

320 **Acknowledgements** The authors would like to thank Dr. Loris Barale, the University of  
321 Turin and all the cats' owners for their help with this study.

322

323 **Conflict of interest** The authors declared no potential conflict of interest with respect to  
324 the research, authorship, and/or publication of this article.

325

326

327

328

329

330

331

332

333 **References**

- 334 1. Merola, I, Mills, DS. **Behavioural signs of pain in cats: an expert consensus.**  
335 *PLoS One* 2016; 11: DOI: 10.1371/journal.pone.0150040.
- 336 2. Merola, I, Mills, DS. **Systematic review of the behavioural assessment of pain**  
337 **in cats.** *J Feline Med Surg* 2016; 18: 60–76.
- 338 3. Reid J, Scott EM, Calvo G et al. **Definitive Glasgow acute pain scale for cats:**  
339 **validation and intervention level.** *Vet Rec* 2017; 180: 449.
- 340 4. Brondani, JT, Mama, KR, Luna, SP. **Validation of the English version of the**  
341 **UNESP-Botucatu multidimensional composite pain scale for assessing**  
342 **postoperative pain in cats.** *BMC Vet Res* 2013; 9: 143.
- 343 5. Holden E, Calvo G, Collins M et al. **Evaluation of facial expression in acute**  
344 **pain in cats.** *J Small Anim Pract* 2014; 55: 615-621.
- 345 6. Grubb T. **What do we really know about the drugs we use to treat chronic**  
346 **pain?** *Top in Companion Anim Med* 2010; 25: 10-19.
- 347 7. Lascelles DB, Dong YH, Marcellin-Little DJ, et al. **Relationship of orthopedic**  
348 **examination, goniometric measurements, and radiographic signs of**  
349 **degenerative joint disease in cats.** *BMC Veterinary Research* 2012; 8:10.
- 350 8. Dimitroulas T, Duarte RV, Behura A et al. **Neuropathic pain in osteoarthritis:**  
351 **a review of pathophysiological mechanisms and implications for treatment.**  
352 *Semin Arthritis Rheum* 2014; 44: 145-154.

- 353 9. Gagnon A, Brown D, Moreau M et al. **Therapeutic response analysis in dogs**  
354 **with naturally occurring osteoarthritis.** *Vet Anaesth Analg* 2017; 44: 1373-  
355 1381.
- 356 10. Knazovicky D, Helgeson ES, Case B et al. **Replicate effects and test-retest**  
357 **reliability of quantitative sensory threshold testing in dogs with and without**  
358 **chronic pain.** *Vet Anaesth Analg* 2017; 44: 615-624.
- 359 11. Benito J, Depuy V, Hardie E, et al. **Reliability and discriminatory testing of a**  
360 **client-based metrology instrument, feline musculoskeletal pain index (FMPI)**  
361 **for the evaluation of degenerative joint disease-associated pain in cats.** *Vet J*  
362 2013; 196: 368-373.
- 363 12. Guillot M, Moreau M, Heit M et al. **Characterization of osteoarthritis in cats**  
364 **and meloxicam efficacy using objective chronic pain evaluation tools.** *Vet J*  
365 2013; 196: 360-367.
- 366 13. Klinck MP, Rialland P, Guillot M et al. **Preliminary validation and reliability**  
367 **testing of the Montreal Instrument for cat arthritis testing, for use by**  
368 **veterinarians, in a colony of laboratory cats.** *Animals (Basel)* 2015; 5: 1252-  
369 1267.
- 370 14. Addison ES and Clements DN. **Repeatability of quantitative sensory testing in**  
371 **healthy cats in a clinical setting with comparison to cats with osteoarthritis.**  
372 *J Feline Med Surg* 2017; 19: 1274-1282.

- 373 15. Stadig S, Lascelles BD, Bergh A. **Do cats with a cranial cruciate ligament**  
374 **injury and osteoarthritis demonstrate a different gait pattern and behaviour**  
375 **compared to sound cats?** *Acta Vet Scand* 2016; 58:71-79.
- 376 16. Kim JS, Ahmadiania K, Li X et al. **Development of an Experimental Animal**  
377 **Model for Lower Back Pain by Percutaneous Injury-Induced Lumbar Facet**  
378 **Joint Osteoarthritis.** *J Cell Physiol* 2015; 230: 2837-2847.
- 379 17. Girard P, Verniers D and Coppé MC. **Nefopam and ketoprofen synergy in**  
380 **rodent models of antinociception.** *Eur J Pharmacol* 2008; 584: 263-271.
- 381 18. Reynoso-Moreno I, Najjar-Guerrero I, Escareño N, et al. **An endocannabinoid**  
382 **uptake inhibitor from black pepper exerts pronounced anti-inflammatory**  
383 **effects in mice.** *J Agric Food Chem* 2017; 65: 9435-9442.
- 384 19. Walter SD, Eliasziw M, Donner A. **Sample size and optimal designs for**  
385 **reliability studies.** *Stat Med* 1998; 17:101-110.
- 386 20. LaFlamme DP. **Development and validation of a body condition score system**  
387 **for cats: a clinical tool.** *Feline Pract* 1997; 25: 13-18.
- 388 21. Nie H, Arendt-Nielsen L and Andersen H. Temporal summation of pain evoked  
389 by mechanical stimulation in deep and superficial tissue. *J Pain* 2005; 6: 348-355.
- 390 22. Bland JM and Altman DG. **Statistical methods for assessing agreement**  
391 **between two methods of clinical measurement.** *Lancet* 1986; 327: 307–310.

- 392 23. Cicchetti DV. **Guidelines, criteria, and rules of thumb for evaluating normed**  
393 **and standardized assessment instruments in psychology.** *Psychological*  
394 *Assessment* 1994; 6: 284–290.
- 395 24. Clarke SP, Mellor D, Clements DN, et al. **Prevalence of radiographic signs of**  
396 **degenerative joint disease in a hospital population of cats.** *Vet Rec* 2005; 157:  
397 793-799.
- 398 25. Lascelles BD, Henry JB, Brown J, et al. **Cross-sectional study of the prevalence**  
399 **of radiographic degenerative joint disease in domesticated cats.** *Vet Surg*  
400 2010; 39: 535-544.
- 401 26. Keller GG, Reed AL, Lattimer JC, et al. **Hip dysplasia: a feline population**  
402 **study.** *Vet Radiol Ultrasound* 1999; 40: 460-440.
- 403 27. Hardie EM, Roe SC, Martin FR. **Radiographic evidence of degenerative joint**  
404 **disease in geriatric cats: 100 cases (1994-1997).** *J Am Vet Med Assoc* 2002; 220:  
405 628-632.